## Subject: Reply of CUH on the quarries raised by the bidders during the Pre-bid Meeting held on 09/10/2025.

Name of Equipment: Mercury Intrusion Machine

Sr.	Name of	CUH Specification as per NIT	Bidder's queries	Reply of CUH
No.	firm	Coll Specification as per 1411	bluder 5 queries	Reply of Coll
A.	Anton Paar India Pvt. Ltd.	Analysis Stations (Point No. 2(iii) of technical specifications):  High pressure station – 1 no - 0-33,000 psia	Point no- 2 In Analysis Stations: High pressure station – 1 no – 20 - 33,000 psia instead of 0- 33,000 psia	The Point No. 2(iii) may be read as under:  "High pressure station – 1 no - 20-33,000 psia"
		Additional Accessories (Point No.4 (8) of technical specifications):  Also should include hose for connection to existing fume hood for safe removal of mercury vapor	Point no- 4 - In additional accessories – Please mentioned Fume hood is required by supplier or Provided by CUH.	Fume hood is required to be supplied by supplier along with all accessories for safe removal of mercury vapor