

Subject: Reply of CUH on the queries raised by the bidders during the Pre-bid Meeting held on 09/02/2026.

Name of Equipment: Time Correlated Single Photon Counting (TCSPC) with low temperature photoluminescence (PL) spectroscopy for the Department of Physics & Astrophysics.

S r. No	Name of firm	Reference to NIT No. CUH/E&GA/HEFA/CIC /2025-26/34	Bidder's comment/observations	Reply from CUH/Technic al Committee
1.	Horiba India Pvt. Ltd.	<p>Excitation Sources As per tender Documents 450 W Xenon Lamp</p> <ul style="list-style-type: none"> • 450 W xenon lamp with integrated power supply • Wavelength range- 230 nm -900 nm • display of power, voltage, current, time elapsed <p>60 W Microsecond Flashlamp</p> <ul style="list-style-type: none"> • Pulsed high energy xenon flashlamp • repetition frequency under computer control • optical pulse width 1.5 μs to 2.5 μs • average power up to 60 W • repetition rate: 0.1 Hz – 100 Hz <p>Wavelength range- 250 nm -850 nm</p>	<p>Excitation Sources We Request to please incorp 450W or 75 W and 60W or 15W Plused Lamp for phosphorescence as a source and enable us to participate this bid. Our unique design with 75W providing the high intensity, stable light and broad spectral coverage source necessary for precise fluorescence measurements.</p>	450W Xenon lamp and Microsecond Flash Lamp (60W or 15W) for phosphorescence
		<p>TCSPC Lifetime Measurements and Electronics</p> <p>510 nm (\pm 5nm) Picosecond Diode Laser pulse width ~85 ps or better with rep Rate upto 20MHz or better</p>	<p>TCSPC Lifetime Measurements and Electronics</p> <p>We request to please allow us to bid with 510 nm (\pm 10nm) Picosecond Diode Laser pulse width ~125 ps or better with rep Rate upto 25 MH</p>	No Change



		(As per Tender Documents)		
2.	Specialise Instruments Marketing Company	EMD (As per Tender Documents)	1. EMD: As we have already submitted the EMD of the same amount through RTGS as per details given in tender, for the previous tender refer no.(CUH/E&GA/HEFA/CIC /2025-26/12,which got cancelled),we hereby request you to kindly consider the same.	Yes, you can use the same EMD
		2. Payments (As per Tender Documents)	Payment terms: 90% against delivery and 10% after installation	No Change
		3. Delivery (As per Tender Documents)	Delivery 4-5 months	No Change

