

Subject: Reply of CUH on the queries raised by the bidders during the Pre-bid Meeting held on 23/02/2026.

Name of Equipment: Probe Station with Semiconductor Characterization for the Department of Physics & Astrophysics.

Sr. No.	Name of firm	Reference to NIT No. CUH/E&GA/HEFA/CIC/2025-26/41	Bidder's comment/observations	Reply from CUH/Technical Committee
1.	Silicom Electronics Pvt. Ltd.	<p>1) Chapter 3: Schedule Of Requirements</p> <p>i. Site preparation: The bidder shall be fully responsible for complete site preparation as per equipment specifications, including any civil, electrical, or environmental requirements before delivery. The University shall provide only the required space as per the size requirement furnished with simple tile flooring and provision of Air conditioning in the building. A three-phase electricity point along with earthing arrangements shall be provided in that space. The rest of the wiring, MCB installations, base isolation etc. as per the equipment requirement shall be responsibility of the vendor for successful installation and testing &amp; commissioning.</p>	<p>As the equipment required will be having Single Phase inputs. We request you to provide single phase sockets &amp; switches. Multi sockets with Power (16A) &amp; 6A current will be preferred. As these are table top units, Tables will be required for the setup along with nearby or on table provision of switch sockets.</p>	Agreed
		<p>2.</p> <p>viii. The equipment shall be supplied with a UPS (5 kVA) along with batteries for a backup time of one hour.</p>	<p>As UPS is generally required for power backup, One hour backup time will be too high &amp; requires external batteries. We suggest the backup time should be reduced to 15-20 min for power backup with Internal Maintenance free battery.</p>	Agreed

		<p><b>3. 04 (ii)</b> The complete system must function as an integrated platform, with centralized control and data acquisition enabled via the SPA. A unified or seamlessly integrated software interface must be provided.</p>	<p>Integration will be provided between the products compatible &amp; required.</p>	<p><b>Agreed</b></p>
2.	<p><b>Agmatel India Pvt. Ltd.</b></p>	<p>The representative attended the pre-bid meeting and participated in discussions. However, no written query was submitted.</p>		